



## Surprisingly low frequency attenuation effects in long tubes when measuring turbulent fluxes at tall towers

Ibrom, Andreas; Brændholt, Andreas; Pilegaard, Kim

*Published in:*  
Geophysical Research Abstracts

*Publication date:*  
2016

*Document Version*  
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

*Citation (APA):*  
Ibrom, A., Brændholt, A., & Pilegaard, K. (2016). Surprisingly low frequency attenuation effects in long tubes when measuring turbulent fluxes at tall towers. *Geophysical Research Abstracts*, 18, [EGU2016-12800-1].

---

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

## **Surprisingly low frequency attenuation effects in long tubes when measuring turbulent fluxes at tall towers**

Andreas Ibrom, Andreas Brændholt, and Kim Pilegaard

Technical University of Denmark (DTU), Department of Environmental Engineering, Lyngby, Denmark (anib@env.dtu.dk)

The eddy covariance technique relies on the fast and accurate measurement of gas concentration fluctuations. While for some gasses robust and compact sensors are available, measurement of, e.g., non CO<sub>2</sub> greenhouse gas fluxes is often performed with sensitive equipment that cannot be run on a tower without massively disturbing the wind field. To measure CO and N<sub>2</sub>O fluxes, we installed an eddy covariance system at a 125 m mast, where the gas analyser was kept in a laboratory close to the tower and the sampling was performed using a 150 m long tube with a gas intake at 96 m height.

We investigated the frequency attenuation and the time lag of the N<sub>2</sub>O and CO concentration measurements with a concentration step experiment. The results showed surprisingly high cut-off frequencies (close to 2 Hz) and small low-pass filter induced time lags (< 0.3 s), which were similar for CO and N<sub>2</sub>O. The results indicate that the concentration signal was hardly biased during the ca 10 s travel through the tube. Due to the larger turbulence time scales at large measurement heights the low-pass correction was for the majority of the measurements < 5%. For water vapour the tube attenuation was massive, which had, however, a positive effect by reducing both the water vapour dilution correction and the cross sensitivity effects on the N<sub>2</sub>O and CO flux measurements.

Here we present the set-up of the concentration step change experiment and its results and compare them with recently developed theories for the behaviour of gases in turbulent tube flows.